



# HIGHLY PURIFIED ARGON

GRADE N-5.0

## COMPOSITION

ARGON

## SPECIFICATION

Min. 99.999%

## IMPURITIES

OXYGEN	< 2.0 ppm
MOISTURE	< 3.0 ppm
THC	< 1.0 ppm*
NITROGEN	< 4.0 ppm

A certificate of conformity is issued with every cylinder.

\* THC testing of one cylinder is carried per batch.

<b>General Information</b>	
<b>Valve Outlet</b>	BS 341
<b>Cylinder size</b>	No-3
<b>Contents m<sup>3</sup></b>	J
<b>Pressure PSI at 15°C</b>	6.8 m3
	1980 ±
	5%

\_\_\_\_\_  
Analytical Chemist

\_\_\_\_\_  
Review By



Pakistan Oxygen Limited (Previously Linde  
Pakistan Limited)

PO Box # 4845, Dockyard Road, West Wharf Karachi-  
74000, Pakistan.

[www.pakxygen.com](http://www.pakxygen.com)