



HIGHLY PURIFIED NITROGEN

GRADE N-5.0 Specification

COMPOSITION

NITROGEN

SPECIFICATION

Min 99.999%

IMPURITIES

OXYGEN

< 3.0 ppm

MOISTURE

< 3.0 ppm

THC

< 1.0 ppm *

CO/CO₂

< 0.5 ppm **

A certificate of conformity is issued with every cylinder.

All cylinders are checked 100%.

*This test is carried out on one cylinder per batch.

** This can be done on customer request.

General Information

Valve Outlet

BS 341 No-3

Cylinder size

J

Contents m³

6.2 m³

Pressure PSI at 15°C

1980 ± 5%

Pakistan Oxygen Limited (Previously Linde Pakistan Limited)
PO Box # 4845, Dockyard Road, West Wharf Karachi-74000, Pakistan.
www.pakoxygen.com